



All correspondence should be addressed to the Director General

LET_LSE_PMS_0013/09/25
In reply, please quote

ZAMBIA MEDICINES REGULATORY AUTHORITY

PUBLIC NOTICE

Date: 16th September 2025

To: General Public

URGENT PUBLIC ALERT:

RECALL OF LATEX EXAMINATION GLOVES PRE-POWDERD BATCH NO. S 4B001 & 4B003 MANUFACTURED BY TOP GLOVES SDN. BHD MALAYSIA

FOR IMMEDIATE RELEASE – The Zambia Medicines Regulatory Authority (ZAMRA) is a statutory body established under an Act of Parliament, the Medicines and Allied Substances Act No. 3 of 2013 of the Laws of Zambia. The main mandate of ZAMRA is to ensure the quality, safety and efficacy of medicines and allied substances for human and animal health protection.

On this background, the Authority wishes to notify all healthcare professionals, health facilities, pharmaceutical outlets and members of the public that it has instructed Lumumba Pharmaceuticals Limited to urgently **RECALL** the above-stated batches of due to failure to meet its Water Tightness (Leak) specifications.

Product details:

S/N	Name of Product	Manufacturer	Batch No.	Expiry Date	Laboratory Test Report	Test Results
1.	Latex Examination Gloves, Pre-Powdered	Top Gloves SDN, BHD, Malaysia	4B003	11/2029	2025/07/TL57796	Non-compliant
			4B001	03/2029	2024/11/TL52716	Non-compliant

RISK

Gloves with failed leak test compromise barrier protection and pose serious risks, including exposure to infectious agents or hazardous chemicals, cross-contamination, and loss of sterility.

If you are in possession of these batches, please return them to your healthcare provider or supplier. Pharmaceutical outlets and health facilities in possession of the affected batches should quarantine the products and return them to their supplier(s), for collection and disposal by the importer.

Should you need further clarification, please do not hesitate to contact our Secretariat.

Makomani Siyanga (Mr)
DIRECTOR-GENERAL

Head Office
Plot No: 2350/M
Off Kenneth Kaunda International Airport Road.
P.O. Box 31890, Lusaka, ZAMBIA
Tel: +260 211 432 350 /432 351
E-mail: pharmacy@zamra.co.zm

Report Adverse Reactions to:
Pharmacovigilance Unit, Lusaka
Tel: +260 211 432 356
E-mail: npvu@zamra.co.zm
Website: www.zamra.co.zm